



財團法人全國認證基金會  
Taiwan Accreditation Foundation

## Certificate of Accreditation

(Certificate No : LN0881-240923)

This is to certify that

**National Measurement Laboratory, R. O. C.**  
**National Measurement Laboratory, R. O. C.(Temperature, Humidity)**  
321, Sec. 2, Kuang Fu Rd., Hsinchu, Taiwan, R.O.C.

**is accredited in respect of laboratory**

**Accreditation Criteria** : ISO/IEC 17025:2017 ; CNS 17025:2018

**Accreditation Number** : N0881

**Originally Accredited** : May 15, 2002

**Effective Period** : May 15, 2021 to May 14, 2026

**Accredited Scope** : Calibration Field, see described in the Appendix

**Specific Accreditation Program** : Accreditation Program for National Metrology Institutes



Scan to verify

*Yi-Ling Chen*

Yi-Ling Chen  
President, Taiwan Accreditation Foundation  
September 23, 2024

Accreditation Number : N0881

Laboratory Head : LAN, Yu-Ping

## Temperature/Humidity

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units		explanation	value
KE1002 Platinum resistance thermometer, Fixed-point Cell	FLUKE/5960A, ISOTECH/17924, FLUKE/5900E, Hart Scientific/5901A, Hart Scientific/5943, ISOTECH/17668, FLUKE/5904, NML/Sn, ISOTECH/17669, ISOTECH/17671, NML/Al, FLUKE/5907, FLUKE/5908, ISOTECH/17673	Instrument Calibration Technique for the Fixed-Point Calibration of Platinum Resistance Thermometer (Document No.: 07-3-82-0061)	-190	°C	962	°C	FP of Ag	6.0	mK
			-190	°C	962	°C	FP of Al	4.1	mK
			-190	°C	962	°C	FP of Zn	2.1	mK
			-190	°C	962	°C	FP of Sn	0.52	mK
			-190	°C	962	°C	FP of In	0.85	mK
			-190	°C	962	°C	MP of Ga	0.37	mK
			-190	°C	962	°C	TP of H <sub>2</sub> O	0.16	mK
			-190	°C	962	°C	TP of Hg	0.43	mK
-190	°C	962	°C	TP of Ar	0.81	mK			
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units		explanation	value
KE1002 Resistance thermometer detector, Resistance thermometer, Thermometer containing display meter (Digital thermometer)	ISOTECH 909/25.5, ROSEMOUNT 162CE	Instrument Calibration Technique for Resistance Temperature Thermometer (Document No.: 07-3-82-0060)	0	°C	0	°C		6.3 (0.0063)	mK (°C)
			5	°C	10	°C		6.8 (0.0068)	mK (°C)
			>10	°C	20	°C		7.0 (0.0070)	mK (°C)
			>20	°C	30	°C		7.0 (0.0070)	mK (°C)
			>30	°C	40	°C		7.4 (0.0074)	mK (°C)
			>40	°C	50	°C		6.9 (0.0069)	mK (°C)
			>50	°C	60	°C		7.0 (0.0070)	mK (°C)
			>60	°C	70	°C		7.3 (0.0073)	mK (°C)
			>70	°C	80	°C		8.0 (0.0080)	mK (°C)
			>80	°C	100	°C		11 (0.011)	mK (°C)
			>100	°C	150	°C		8.7 (0.0087)	mK (°C)
			>150	°C	200	°C		9.0 (0.0090)	mK (°C)
			>200	°C	250	°C		11 (0.011)	mK (°C)
>250	°C	300	°C		18 (0.018)	mK (°C)			
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									
KE1004 type R, S, B thermocouple	Hart -5943-09, CMS-Sn-1-05, Fluke-5906-Zn-06116, Fluke-5907-Al-07170, Fluke-5908-Ag-08077	Instrument Calibration Technique for the Fixed-Point Calibration of Noble Metal Thermocouples (Document No.: 07-3-89-0009)	29.7646	°C	29.7646	°C	MP of Ga	0.20	°C
			231.928	°C	231.928	°C	FP of Sn	0.14	°C
			419.527	°C	419.527	°C	FP of Zn	0.13	°C
			660.323	°C	660.323	°C	FP of Al	0.12	°C
			961.78	°C	961.78	°C	FP of Ag	0.11	°C
			0	°C	961.78	°C		0.20	°C
>961.78	°C	1200	°C		0.41	°C			
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units		explanation	value
KE1007 Standard radiation thermometer	Ag: CHINO IR-RO-AG, Cu: CHINO IR-RO-Cu	Instrument Calibration Technique for the Fixed-Point Calibration of Radiation Thermometers (Document No.: 07-3-90-0113)	800	°C	800	°C	650 nm	0.38	°C
			900	°C	900	°C	650 nm	0.30	°C
			1000	°C	1000	°C	650 nm	0.31	°C
			1100	°C	1100	°C	650 nm	0.35	°C
			1200	°C	1200	°C	650 nm	0.37	°C
			1300	°C	1300	°C	650 nm	0.45	°C
			1400	°C	1400	°C	650 nm	0.66	°C
			1500	°C	1500	°C	650 nm	0.89	°C
			1600	°C	1600	°C	650 nm	1.2	°C
			1700	°C	1700	°C	650 nm	1.5	°C
			1800	°C	1800	°C	650 nm	1.8	°C
			1900	°C	1900	°C	650 nm	2.1	°C
			2000	°C	2000	°C	650 nm	2.5	°C
			800	°C	800	°C	900 nm	0.28	°C
			900	°C	900	°C	900 nm	0.24	°C
			1000	°C	1000	°C	900 nm	0.26	°C
			1100	°C	1100	°C	900 nm	0.31	°C
			1200	°C	1200	°C	900 nm	0.34	°C
			1300	°C	1300	°C	900 nm	0.41	°C
			1400	°C	1400	°C	900 nm	0.55	°C
			1500	°C	1500	°C	900 nm	0.71	°C
			1600	°C	1600	°C	900 nm	0.89	°C
			1700	°C	1700	°C	900 nm	1.1	°C
			1800	°C	1800	°C	900 nm	1.4	°C
1900	°C	1900	°C	900 nm	1.6	°C			
2000	°C	2000	°C	900 nm	1.9	°C			
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									





calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model		document name /no.	minimum value	units	maximum value		units	explanation
KE1007 Linear Pyrometer	Ag: CHINO IR-RO-AG, Cu: CHINO IR-RO-Cu	Instrument Calibration Technique for the Fixed-Point Calibration of Radiation Thermometers (Document No.: 07-3-90-0113)	961.78	°C	961.78	°C	(@650 nm) -Ag	0.25	°C
			1084.62	°C	1084.62	°C	(@650 nm) -Cu	0.33	°C
			961.78	°C	961.78	°C	(@900 nm) -Ag	0.25	°C
			1084.62	°C	1084.62	°C	(@900 nm) -Cu	0.30	°C
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									
KE1007 Radiation thermometer	ASL/T100-250-ID	Instrument Calibration Technique for the Comparative Calibration of Room Temperature Radiation Thermometer (Document No.: 07-3-97-0003)	10	°C	90	°C		0.1	°C
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									
KE2001 Hygrometer electronic psychrometer, humidity transducer, digital humidity analyzer,	Thunder Scientific 2500ST	Instrument Calibration Technique for Two- Pressure Humidity Generator (2500) Calibration System (Document No.: 07-3-99-5145)	10	%	98	%	relative humidity	(0.0038×RH +0.0306)	%RH
			0	°C	<35	°C	Temperature	0.064	°C
			35	°C	<69.5	°C	Temperature	0.12	°C
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE2003 Optical chilled dew point hygrometer	Thunder Scientific 2500ST	Instrument Calibration Technique for Two-Pressure Humidity Generator (2500) Calibration System (Document No.: 07-3-99-5145)	-27	°C	25	°C	Dew point	0.068	°C
			>25	°C	<68	°C	Dew point	0.089	°C
Approval Signatory: KO, Hsin-Yi; YEH, Chien-Chih; TSAI, Shu-Fei; TU, Tsung-Hsien									

Note : Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.  
(Null Below)

